



## Guidance

# Register as a manufacturer to sell medical devices

From: [Medicines and Healthcare products Regulatory Agency](#)  
Part of: [Patient safety](#) and [Medical devices regulation and safety](#)  
First published: 18 December 2014  
Last updated: 1 April 2017, [see all updates](#)

Register a manufacturer or authorised representative (agent) for a manufacturer, to sell medical devices, including in vitro diagnostic (IVD) medical devices.

## Contents

- [Overview](#)
- [When you must register](#)
- [Eligible devices and their codes](#)

## Overview

If you place certain medical devices on the EU market you or your authorised representative (agent) must be registered with the competent authority (national health regulator) in the EU state where you have an office or place of business. In the UK,

- [CE marking](#)
- [Apply to register with the Device Online Registration System \(DORS\)](#)
- [Making changes to your DORS registration](#)
- [Contact](#)

Medicines and Healthcare Products Agency (MHRA) is the competent authority for the registration of medical devices. MHRA will only register manufacturers or authorised representatives that have a place of business in the UK.

Before you register, you must first [complete a conformity assessment](#) , so you can put the CE marking on your products.

## When you must register

You must inform a competent authority (in the UK this is the MHRA) when you first place your device on the market. You must register if you or your company sells, leases, lends or gifts:

- class I devices you have manufactured
- class I devices you have refurbished or re-labelled with your own name
- any system or procedure pack containing at least one medical device
- custom-made devices
- [‘in vitro diagnostic medical device’ \(IVD\)](#) you have manufactured
- in vitro diagnostic medical devices (IVDs) undergoing performance evaluation

Note: We do not register devices classified as class IIa, IIb, III or active implantable devices. For these, you need to follow the appropriate [conformity assessment route](#), which includes being assessed by a notified body.

Manufacturers without a place of business in the EU need to appoint an authorised representative in the EU. Only one authorised representative can be designated within the EU for each product type.

## Eligible devices and their codes

## **Class I devices**

Examples of class I devices include:

- dental and surgical instruments
- stethoscopes and ophthalmoscopes
- bandages and splints
- spectacle lenses and frames
- treatment chairs and hospital beds

Examples of products that are not medical devices include:

- toiletries
- tattooing instruments
- protective equipment
- swabs, wipes and disinfectants

The European Commission has a guidance document [Definitions of “medical devices”, “accessory” and “manufacturer” \(MEDDEV 2.1/1\)](#).

## **Custom-made devices, systems and procedure packs**

You will need to register if your company:

- places devices bearing the CE marking on the market, under your own name in a system or a procedure pack within their intended purposes and within the limits of uses specified by their original manufacturers
- sterilises, for the purpose of placing on the market under your own name, systems or procedure packs or other CE marked medical devices designed by the manufacturer to be sterilised before use

There is more information about custom-made devices and examples of the information we need on your statement on [this page](#).

To register your medical device with the MHRA, you can use either of the following lists of codes:

- [MHRA generic device description codes](#)
- the [Global Medical Devices Nomenclature System](#)

### **In vitro diagnostic medical devices**

Examples of in vitro diagnostic (IVD) medical devices include cultures, reagents, antigens, serums, plasmas etc and testing kits and equipment for these.

You will need to specify the correct codes for your IVD medical devices.

You can use either of the following lists of codes to register your IVD device with the MHRA:

- the [Global Medical Devices Nomenclature System](#)
- the [European Diagnostic Manufacturers Association \(EDMA\)](#)

### **CE marking**

The CE marking is a declaration of conformity with the EU's Medical Devices Directive (MDD).

For class I devices (those with the lowest risk level) you must ensure that your product complies with all the relevant essential requirements of the MDD and draw up a written statement to this effect (self-declaration).

If you manufacture sterile products or devices with a measuring function you must apply to a notified body for certification.

Once you are satisfied that your products meet all the relevant essential requirements, you can register with MHRA (or other EU authority), then you can put the CE mark on the product(s) and place them on the market.

Find out more about the conformity assessment for medical devices in section 2.5 of Guidance MEDDEVs on the [European Commissions website](#).

## **Apply to register with the Device Online Registration System (DORS)**

Using DORS you can register medical devices and IVD devices and manage your account details. First you need to create an account on DORS and wait for the email from us confirming that it is active before you can start registering your products.

### **Increase to the Devices Registration Fee**

Please note as of 1st April 2017 The Medical Devices (Fees Amendment) Regulations 2017 will come into force. The amending Regulation includes an increase to the registration statutory fee from £70.00 to £100.00 per notification of medical and/or in vitro diagnostic device(s) being placed on the UK market.

You should note payment is still required for new and subsequent changes to the registered details held, and should accompany the notification upon submission to the UK Competent Authority.

Start the process on the [device online registration system \(DORS\)](#) and we will email you when it's activated.

You must pay £100 via DORS per registration.

There is more information about DORS [here](#).

If you are an authorised representative (AR) of a non-EU manufacturer you must provide written evidence that you are acting with the consent of the manufacturer. We need an official letter, on headed paper, from the manufacturer stating your

company's name and address – this is the designation letter.

If you are taking over the role from an existing authorised representative we need the designation letter and a copy of the letter cancelling the service with the previous representative.

The European Commission has [guidance for authorised representatives](#).

## **Making changes to your DORS registration**

You must inform MHRA, by updating your DORS record, if you want to change:

- address
- addition of device types
- manufacturer name
- authorised representative
- status of an IVD, for example a change from 'performance evaluation' to 'new'

You must pay £100 for each change request but you can change more than one detail within each registration request. We have to check these changes so they won't appear immediately in your DORS record.

Changes you can make for free by updating your DORS record include:

- removing a device from your records
- changing a contact person
- changing a telephone number and/or email address.

There is no need to tell us if you decide to sell extra products that fall under the same generic product code you have already registered with us.

## **Register of manufacturers**

Once registered, your company will be added to the [Public Access Database for Medical Device Registration](#). Records are listed by manufacturer and device and include contact details.

Manufacturers of in vitro diagnostic medical devices will not be published on this database, as the IVD Directive confidentiality clause still applies.

Registering your devices with MHRA does not mean that we give you any form of accreditation, certification or approval for the device. It is just a notification of your activities.

## Contact

If you are already registered with us, please contact us under the following alphabetical splits for your company name or the surname under which the registration is held, quoting your reference number:

- A – K Nashay MODI-Wilson on 020 3080 7253
- L – Z Barbra Clarke on 020 3080 7318

Email: [Devices.regulatory@mhra.gsi.gov.uk](mailto:Devices.regulatory@mhra.gsi.gov.uk)

Published:

**18 December 2014**

Updated:

**1 April 2017**

[+ full page history](#)

From:

**Medicines and Healthcare products Regulatory Agency**

Part of:

**Patient safety**

**Medical devices regulation and safety**

[Is there anything wrong with this page?](#)

## Services and information

---

[Benefits](#)

[Births, deaths, marriages and care](#)

[Business and self-employed](#)

[Childcare and parenting](#)

[Citizenship and living in the UK](#)

[Crime, justice and the law](#)

[Disabled people](#)

[Driving and transport](#)

[Education and learning](#)

[Employing people](#)

[Environment and countryside](#)

[Housing and local services](#)

[Money and tax](#)

[Passports, travel and living abroad](#)

[Visas and immigration](#)

[Working, jobs and pensions](#)

## Departments and policy

---

[How government works](#)

[Departments](#)

[Worldwide](#)

[Policies](#)

[Publications](#)

[Announcements](#)

---

[Help](#) [Cookies](#) [Contact](#) [Terms and conditions](#)

[Rhestr o Wasanaethau Cymraeg](#) Built by the [Government Digital Service](#)

**OGI** All content is available under the [Open Government Licence v3.0](#), except where otherwise stated



© Crown copyright