

Contains Nonbinding Recommendations

Draft – Not for Implementation

1 **Expansion of the Abbreviated 510(k)**
2 **Program: Demonstrating Substantial**
3 **Equivalence through Performance**
4 **Criteria**

6 **Draft Guidance for Industry and**
7 **Food and Drug Administration**

8 ***DRAFT GUIDANCE***

9 **This draft guidance document is being distributed for comment purposes only.**

10 **Document issued on April 12, 2018.**

11 You should submit comments and suggestions regarding this draft document within 90 days of
12 publication in the *Federal Register* of the notice announcing the availability of the draft
13 guidance. Submit electronic comments to <https://www.regulations.gov>. Submit written
14 comments to the Dockets Management Staff (HFA-305), Food and Drug Administration, 5630
15 Fishers Lane, rm. 1061, Rockville, MD 20852. Identify all comments with the docket number
16 listed in the notice of availability that publishes in the *Federal Register*.

17 For questions about this document regarding CDRH-regulated devices, contact the 510(k) Staff at
18 301-796-5640. For questions about this document regarding CBER-regulated devices, contact the
19 Office of Communication, Outreach, and Development (OCOD) at 1-800-835-4709 or 240-402-
20 8010.



33 U.S. Department of Health and Human Services
34 Food and Drug Administration
35 Center for Devices and Radiological Health
36 Center for Biologics Evaluation and Research

Preface

38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55

Additional Copies

CDRH

Additional copies are available from the Internet. You may also send an e-mail request to CDRH-Guidance@fda.hhs.gov to receive a copy of the guidance. Please use the document number 17038 to identify the guidance you are requesting.

CBER

Additional copies are available from the Center for Biologics Evaluation and Research (CBER), Office of Communication, Outreach, and Development (OCOD), 10903 New Hampshire Ave., Bldg. 71, Room 3128, Silver Spring, MD 20993-0002, or by calling 1-800-835-4709 or 240-402-8010, by email, ocod@fda.hhs.gov or from the Internet at <https://www.fda.gov/BiologicsBloodVaccines/GuidanceComplianceRegulatoryInformation/Guidances/default.htm>.

Table of Contents

56
57
58 I. Introduction..... 4
59 II. Background..... 4
60 III. Policy..... 6
61 A. Intended use and technological characteristics 7
62 B. Identification of performance criteria 8
63 C. FDA review of data..... 8
64 D. Modifications to the list 9
65 Appendix..... 10
66
67

DRAFT

Expansion of the Abbreviated 510(k) Program: Demonstrating Substantial Equivalence through Performance Criteria

Draft Guidance for Industry and Food and Drug Administration Staff

This draft guidance, when finalized, will represent the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff or Office responsible for this guidance as listed on the title page.

I. Introduction

This guidance provides FDA’s current thinking on expanding the use of the Abbreviated 510(k) program for demonstrating substantial equivalence for premarket notification (510(k)) submissions. The intent of the guidance is to describe an optional pathway for certain, well understood device types, where a submitter would demonstrate that a new device meets FDA-identified performance criteria to demonstrate that the device is as safe and effective as a legally marketed device.

FDA's guidance documents, including this draft guidance, do not establish legally enforceable responsibilities. Instead, guidances describe the Agency’s current thinking on a topic and should be viewed only as recommendations, unless specific regulatory or statutory requirements are cited. The use of the word *should* in Agency guidance means that something is suggested or recommended, but not required.

II. Background

For the purposes of determining substantial equivalence, section 513(i)(1)(A) of the Federal Food, Drug, and Cosmetic Act (FD&C Act) provides that:

Contains Nonbinding Recommendations

Draft – Not for Implementation

100
101 [W]ith respect to a device being compared to a predicate device, that device has the same
102 intended use as the predicate device and that the Secretary by order has found that the
103 device –

104 (i) has the same technological characteristics as the predicate device, or

105 (ii) –

106 (I) has different technological characteristics and the information submitted that
107 the device is substantially equivalent to the predicate device contains information,
108 including appropriate clinical or scientific data if deemed necessary by the
109 Secretary or a person accredited under section 523, that demonstrates that the
110 device is as safe and effective as a legally marketed device, and

111 (II) does not raise different questions of safety and effectiveness than the
112 predicate device.

113
114 Through guidance, FDA has explained and clarified how it makes substantial equivalence
115 decisions.¹ As described in that guidance, the 510(k) program has undergone a number of
116 statutory changes since its inception, and FDA has adapted its implementation of the program in
117 response to changing statutory requirements and the evolving medical device landscape. For
118 example, FDA established alternative programs for demonstrating substantial equivalence, the
119 Special 510(k) and the Abbreviated 510(k). The Abbreviated 510(k) submission program relies
120 on the use of guidance documents, special controls, and FDA-recognized consensus standards to
121 facilitate 510(k) review.² The current 510(k) program reflects the current statutory framework
122 and FDA’s implementation of that framework through regulation, guidance, and administrative
123 practice.

124
125 Congress has also amended the FD&C Act to add what are known as the “least burdensome”
126 provisions for medical devices,³ some of which are specific to the 510(k) process.⁴ Their general
127 purpose is to ensure FDA requests the minimum information necessary to adequately address a
128 regulatory question or issue through the most efficient manner at the right time.⁵

¹See “The 510(k) Program: Evaluating Substantial Equivalence in Premarket Notifications [510(k)], Guidance for Industry and Food and Drug Administration Staff,” available at <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM284443.pdf>.

²See “The New 510(k) Paradigm: Alternate Approaches to Demonstrating Substantial Equivalence in Premarket Notifications” final guidance, available at <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/ucm080189.pdf>.

³See sections 513(i)(1)(D)(i) – (iii), 513(a)(3)(D)(iii) – (iv), and 515(c)(5)(A) – (D) of the FD&C Act (21 U.S.C. §360c and §360e), established by the Food and Drug Administration Modernization Act of 1997 (FDAMA) (P. L. 105-115).

⁴See section 513(i)(1)(D)(i) – (iii) of the FD&C Act (21 U.S.C. §360c and §360e), established by FDAMA (P. L. 105-115) and amended by Congress through the Food and Drug Administration Safety and Innovation Act (Public Law 112-144) (FDASIA) and the 21st Century Cures Act (Public Law 114-255) (Cures Act).

⁵On December 16, 2017, FDA issued draft guidance on describing its proposed approach regarding the least burdensome provisions as amended by the Cures Act. When finalized, that guidance will describe the guiding principles and recommended approach for FDA staff and industry when applying least burdensome principles. See

Contains Nonbinding Recommendations

Draft – Not for Implementation

129
130 This guidance focuses on the prong of the substantial equivalence analysis that requires a 510(k)
131 submitter to demonstrate that, despite technological differences, its device is as safe and effective
132 as a legally marketed device. FDA recognizes that, in some cases, demonstrating this through
133 direct comparison testing may create burdens for 510(k) submitters that can be avoided. For
134 example, it may be burdensome for submitters to conduct testing against a substantial number of
135 the appropriate predicate devices to demonstrate equivalence for the necessary set of
136 performance and technological characteristics. Sometimes these burdens are necessary in order
137 to determine substantial equivalence. However, consistent with FDA’s mandate in section
138 513(i)(1)(D) of the FD&C Act to consider the least burdensome means of demonstrating
139 substantial equivalence, this guidance expands the potential use of the Abbreviated 510(k)
140 program by explaining how substantial equivalence for certain device types may be
141 demonstrated in a way that is less burdensome but at least as robust. Use of this expanded
142 program may also streamline the review of 510(k) submissions, thereby reducing burdens on the
143 Agency and possibly review times for individual submissions. In addition, this approach could
144 facilitate healthcare professionals and patients making better informed decisions by ensuring that
145 a device cleared through this pathway meets a transparent set of performance criteria. At the
146 same time, this approach satisfies the statutory standard for demonstrating substantial
147 equivalence.
148

149 **III. Policy**

150 Under section 513(i)(1)(A) of the FD&C Act, substantial equivalence is rooted in comparisons
151 between new devices and predicate devices. However, the FD&C Act does not preclude FDA
152 from using performance criteria to facilitate this comparison. If a legally marketed device
153 performs at certain levels relevant to its safety and effectiveness, and a new device meets or
154 exceeds those levels of performance for the same characteristics, FDA could find that the new
155 device is as safe and effective as the legally marketed device. Instead of reviewing data from
156 direct comparison testing between the two devices, FDA could support a finding of substantial
157 equivalence based on data showing the new device meets or exceeds the level of performance of
158 appropriate predicate device(s). Under the approach expanded in this guidance, a submitter
159 could satisfy the requirement to compare its device with a legally marketed device⁶ by, among
160 other things, demonstrating conformance to objective performance criteria established in FDA
161 guidance, FDA-recognized consensus standards, and/or special controls.⁷ In order to identify the
162 specific set of performance criteria appropriate to satisfy a submitter’s comparison to an
163 appropriate predicate for a given device-type, FDA would ensure that those performance criteria

<https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM588914.pdf>

⁶ See 21 CFR 807.87(f).

⁷ A device must comply with any applicable special controls to be within the type to which the special controls apply regardless of which 510(k) pathway is used, see section 513(a)(1)(B) of the FD&C Act, but there may be instances where conformance to special controls, FDA-recognized standards, and/or FDA-established criteria would also be sufficient to demonstrate a device is as safe and effective as a legally marketed device.

Contains Nonbinding Recommendations

Draft – Not for Implementation

164 represent performance that is equivalent to or exceeds the performance of one or more existing,
165 legally marketed devices of that device type, using equivalent test methodology. Thus, by
166 demonstrating that a new device meets the identified performance criteria, a submitter could
167 demonstrate that the new device is at least as safe and effective, as a legally marketed device, in
168 accordance with sections 513(i)(1)(A)(ii)(I) and 513(i)(1)(D) of the FD&C Act. Note that direct
169 comparison of a new device with a legally marketed device would remain available under a
170 Traditional 510(k).

171
172 The policy in this guidance is an expansion of the approach FDA has long applied through the
173 Abbreviated 510(k) program. When submitting an Abbreviated 510(k), a submitter uses
174 conformity to FDA-recognized consensus standards, FDA guidance, and/or special controls to
175 demonstrate some of the performance characteristics necessary to support a finding of substantial
176 equivalence. In the optional program described here, the Expanded Abbreviated 510(k)
177 program, a submitter would use robust versions of those same mechanisms to demonstrate *all* of
178 the performance characteristics necessary to support a finding of substantial equivalence for a
179 given device type.

180
181 FDA believes that use of performance criteria is only appropriate when FDA has determined that
182 (1) the new device has indications for use and technological characteristics that do not raise
183 different questions of safety and effectiveness than the identified predicate, (2) the performance
184 criteria align with the performance of one or more legally marketed devices of the same type as
185 the new device, and (3) the new device meets the performance criteria. All performance criteria
186 for use of the Expanded Abbreviated 510(k) program will be publicized through FDA guidance
187 developed for purposes of this program, which may reference FDA- recognized consensus
188 standards and special controls. If a device cannot rely entirely on performance criteria identified
189 by FDA to demonstrate substantial equivalence for its submission, it is not appropriate for this
190 program; however, we emphasize that the previously established 510(k) programs in which
191 direct performance comparisons against appropriate predicates are conducted, including
192 Traditional, Special, and (non-expanded) Abbreviated 510(k)s, remain available to submitters.

193 **A. Devices Appropriate for the Expanded Abbreviated** 194 **510(k) program: Intended use and technological characteristics**

195 FDA plans to provide information about the types of devices to which the performance criteria
196 would apply in the guidance establishing the performance criteria. Such information may
197 include the relevant product code, appropriate intended uses, appropriate indications for use, and
198 expectations for technological characteristics. In addition, in individual submissions for the
199 Expanded Abbreviated 510(k) program, FDA will continue to require the identification of
200 predicate device(s) for the intended use and technological characteristics prongs of the
201 substantial equivalence analysis. Clarifying the set of devices for which the performance criteria
202 are appropriate in guidance and having submitters identify a predicate of the same device type
203 will help ensure that a new device that utilizes this program has (1) the same intended use as, and
204 (2) technological characteristics that do not raise different questions of safety and effectiveness
205 from the predicate device.

206

Contains Nonbinding Recommendations

Draft – Not for Implementation

207 If you have questions about whether your new device is within a type identified by FDA as
208 appropriate for an Expanded Abbreviated 510(k), specifically (1) whether your new device is
209 within the scope of devices to which the FDA-identified performance criteria are intended to
210 apply, or (2) whether its indications for use or technological characteristics raise different
211 questions of safety and effectiveness than a predicate device, we recommend that you seek
212 feedback through the Q-Submission program on the appropriateness of using the performance
213 criteria.⁸ FDA believes that it will typically be able to make these determinations without
214 reviewing data, as long as the device clearly falls within the types FDA has identified as
215 appropriate for utilizing this program. However, where FDA determines that additional data are
216 necessary to make these determinations, the Agency may, on a case-by-case basis, review that
217 data before determining whether or not the device is appropriate for this Expanded Abbreviated
218 510(k) program.
219

B. Identification of performance criteria

220
221 FDA intends to maintain a list of device types appropriate for the Expanded Abbreviated 510(k)
222 program on the FDA website, accompanied by the guidance documents that identify the
223 performance criteria for each device type, as well as the testing methods recommended in the
224 guidances where feasible, and any other relevant information. These guidance documents may
225 reference consensus standards, or portions of consensus standards, recognized by FDA under
226 section 514 of the FD&C Act, as well as special controls established for that device type. When
227 selecting from established performance criteria and test methodology in standards or guidance
228 and when establishing new performance criteria and test methodology through guidance, FDA
229 intends to rely on the experience and expertise of FDA staff, information in literature, and
230 analyses of data on existing devices within a device type to determine the performance criteria
231 and associated testing methods that could support a finding of substantial equivalence for a given
232 device type. FDA will assure that these criteria represent performance levels that are at least
233 equivalent to the performance of legally marketed devices of the type to which they apply.
234 However, because it is FDA’s responsibility to determine whether a new device reviewed under
235 the 510(k) program is substantially equivalent, the final determination will be FDA’s. We
236 reemphasize that the previously established 510(k) programs in which direct performance
237 comparisons against predicates are conducted, including Traditional, Special, and (non-
238 expanded) Abbreviated 510(k)s, remain available to submitters, as appropriate.

C. FDA review of data

239
240 To support an FDA finding of substantial equivalence through this program, FDA expects a
241 submitter to demonstrate that the new device meets the FDA-identified performance criteria by
242 submitting a declaration of conformity, a summary of the data, and/or underlying data, as
243 appropriate. When the performance criteria and testing methodologies are in an FDA-recognized

⁸ See “Requests for Feedback on Medical Device Submissions: The Pre-Submission Program and Meetings with Food and Drug Administration Staff - Guidance for Industry and Food and Drug Administration Staff,” available at <https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm311176.pdf>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

244 standard and the submitter uses the specified methods to establish that its new device meets the
245 performance criteria, a declaration of conformity should be sufficient. When FDA establishes
246 performance criteria, through guidance and/or special controls, and recommends or specifies
247 testing methodologies, and the submitter uses the method recommended or specified by FDA,
248 submitters should submit a summary of the data. When FDA establishes performance criteria
249 through guidance and/or special controls and recommends or specifies the use of testing
250 methodologies from an FDA-recognized standard, submitters using the recommended or
251 specified testing methodologies should provide a summary of the data in addition to a declaration
252 of conformity.

253

254 FDA may request and review underlying data demonstrating that a new device meets the FDA-
255 identified performance criteria as necessary. When no testing methodology is specified or
256 recommended, or when a submitter chooses to use a testing methodology other than the
257 methodology specified or recommended, submitters should submit underlying data to FDA.

258

259 If data provided by the submitter do not show that the new device meets the performance criteria
260 FDA has identified for the device type, FDA would not be able to find that the new device is
261 substantially equivalent through this program. As previously mentioned, submitters could still
262 use other available 510(k) programs to demonstrate substantial equivalence.

263 **D. Modifications to the list**

264 FDA intends to revise the list of appropriate device types with additional device types and
265 corresponding performance criteria and testing methodology in guidance and on our website over
266 time as appropriate and in accordance with FDA's Good Guidance Practices (21 CFR 10.115).
267 FDA may modify or remove an entry from the list, particularly where new information indicates
268 that the performance criteria in the identified guidance do not fully support a substantial
269 equivalence determination. In such a case, we intend to either remove that device type from the
270 list or note on the list that additional testing may be necessary while the underlying source(s) of
271 the performance criteria are updated. Changes to the list would apply prospectively to devices
272 for which a 510(k) has not yet been submitted.

273

274 **Appendix. Submission Recommendations for an Expanded**
275 **Abbreviated 510(k)**

276 Expanded Abbreviated 510(k)s must comply with the content requirements for premarket
277 notifications submitted in support of substantial equivalence decisions at 21 CFR 807.87. FDA’s
278 [Guidance for Industry and FDA Staff: Format for Traditional and Abbreviated 510\(k\)s](#) (510(k)
279 Format Guidance) provides a general framework for the format and content of an Abbreviated
280 510(k). This appendix provides recommendations on how to apply the recommendations in the
281 general 510(k) Format Guidance to the format and content of an Abbreviated 510(k) that uses the
282 expanded approach described in this guidance. These recommendations are also intended to
283 ensure that the elements recommended in FDA’s guidance “Refuse to Accept Policy for 510(k)s”
284 (RTA Policy Guidance)
285 (<https://www.fda.gov/downloads/medicaldevices/deviceregulationandguidance/guidancedocuments/ucm315014.pdf>)
286 are appropriately included in your submission.
287

288 Consistent with the 510(k) Format Guidance and the RTA Policy Guidance, we recommend that
289 you include the section headings listed, preferably in the sequence outlined below, when
290 submitting an Expanded Abbreviated 510(k) in accordance with this guidance. In some
291 instances, the information in a particular section may not apply to your device. In order to assist
292 review staff, we recommend you retain the section headings in the sequence listed. If you
293 believe a section does not apply, we recommend you include the section and state “This section
294 does not apply” or “N/A” under that heading with a rationale for why the section does not apply.
295 For example, if your device does not contain any software, we recommend you state, “This
296 section is not applicable because the subject device does not contain software” in Section 17
297 titled “Software.”
298

- 299 1. Medical Device User Fee Cover Sheet (Form FDA 3601)
- 300 2. CDRH Premarket Review Submission Cover Sheet (Form FDA 3514)
- 301 3. 510(k) Cover Letter
- 302 4. Indications for Use Statement
- 303 5. 510(k) Summary or 510(k) Statement
- 304 6. Truthful and Accuracy Statement
- 305 7. Class III Summary and Certification
- 306 8. Financial Certification and/or Disclosure Statement (Form FDA 3454)
- 307 9. Certification of Compliance with requirements of ClinicalTrials.gov Data Bank (Form
308 FDA 3674)
- 309 10. Declarations of Conformity and Summary Reports
- 310 11. Executive Summary
- 311 12. Device Description
- 312 13. Substantial Equivalence Discussion
- 313 14. Proposed Labeling
- 314 15. Sterilization and Shelf Life
- 315 16. Biocompatibility
- 316 17. Software

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 317 18. Electromagnetic Compatibility and Electrical Safety
- 318 19. Performance Testing – Bench
- 319 20. Performance Testing – Animal
- 320 21. Performance Testing – Clinical
- 321 22. Other

322

323 Together, the 510(k) Format Guidance and the RTA Policy Guidance provide recommendations
324 for each of these sections. In addition to the recommendations found in the 510(k) Format
325 Guidance, we recommend the following for sections 10, 13, and 14-21:

326

10. Declarations of Conformity and Summary Reports

328

329 To demonstrate your device meets the relevant performance criteria, you should include,
330 depending on the circumstances:

331

- 332 • A declaration of conformity to the standard;⁹
- 333 • a summary report recommended in any relevant device-specific guidance; and/or
- 334 • underlying data demonstrating that the new device meets the FDA-identified performance
335 criteria.

336

337 Which of these approaches is appropriate will depend on the underlying source for the criteria
338 and testing methods, including whether they are contained in an FDA-recognized standard or
339 identified by FDA in guidance and/or special controls, and whether the testing methodology used
340 is recommended or recognized by FDA for this purpose.

341

342 In general, we expect submitters to submit a summary report when criteria are established by
343 guidance and/or special controls. In general, FDA expects submitters to submit a declaration of
344 conformity when using standards. However, if the guidance, special control, or standard does
345 not recommend or specify testing methodologies, FDA expects submitters to submit underlying
346 data.

347

348 It is possible that FDA-identified performance criteria in a guidance or special control would
349 recommend or specify use of an FDA-recognized standard for testing methodology. In this case,
350 FDA would expect submitters to submit a summary report in addition to a declaration of
351 conformity.

352

353 If a submitter chooses not to use FDA-recommended or recognized testing methods, FDA
354 expects the submitter to submit underlying data.

355

356 Finally, FDA may request and review underlying data as necessary.

⁹ See “Required Elements for a Declaration of Conformity to a Recognized Standard (Screening Checklist for All
Premarket Notification (510(k)) Submissions)” available at
<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PremarketSubmissions/PremarketNotification510k/ucm142706.htm>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

357
358
359
360
361
362
363
364
365
366
367
368
369
370
371
372
373
374
375
376
377
378
379
380
381
382
383

13. Substantial Equivalence Discussion

In the substantial equivalence section, we continue to recommend that you identify the predicate by providing its trade name, model number, name of the 510(k) submitter/holder, and 510(k) number, if available.

We recommend that you provide a comparison between your device and the predicate in terms of indications for use and technology.

If you choose to use the Expanded Abbreviated 510(k) program, we do not expect you to provide direct comparison testing against a legally marketed device for performance specifications. Any testing you conduct in accordance with standards or guidance should be as described in sections 10, 13, and 14-21, as applicable.

14 - 21. Proposed Labeling, Sterilization and Shelf Life, Biocompatibility, Software, Electromagnetic Compatibility and Electrical Safety, and Performance Testing – Bench, Animal, and Clinical, as applicable

We continue to recommend that submitters of Abbreviated 510(k)s through this expanded program provide the information described in the 510(k) Format Guidance for the sections on Proposed Labeling, Sterilization and Shelf Life, Biocompatibility, Software, Electromagnetic Compatibility and Electrical Safety, and Performance Testing, except that FDA would not expect your information to describe direct comparison testing against the predicate device. Instead, FDA recommends that you include a declaration of conformity, summary of the data, and/or underlying data, as applicable, demonstrating the new device meets the performance criteria using appropriate testing methods.