

1 **FDA Categorization of Investigational**  
2 **Device Exemption (IDE) Devices to**  
3 **Assist the Centers for Medicare and**  
4 **Medicaid Services (CMS) with**  
5 **Coverage Decisions**

---

8 **Draft Guidance for Sponsors, Clinical**  
9 **Investigators, Industry, Institutional**  
10 **Review Boards and Food and Drug**  
11 **Administration Staff**

12 ***DRAFT GUIDANCE***

13 **This draft guidance document is being distributed for comment purposes only.**

14 **Document issued on June 1, 2016.**

15 You should submit comments and suggestions regarding this draft document within 60 days of  
16 publication in the *Federal Register* of the notice announcing the availability of the draft  
17 guidance. Submit electronic comments to <http://www.regulations.gov>. Submit written  
18 comments to the Division of Dockets Management (HFA-305), Food and Drug Administration,  
19 5630 Fishers Lane, rm. 1061, Rockville, MD 20852. Identify all comments with the docket  
20 number listed in the notice of availability that publishes in the *Federal Register*.

21 For questions about this document, contact Program Operations Staff at 301-796-5640, Center for  
22 Devices and Radiological Health, Food and Drug Administration, 10903 New Hampshire Ave.,  
23 Bldg. 66, rm. 1522, Silver Spring, MD 20993-0002, 301-796-5640. For questions regarding this  
24 document as applied to devices regulated by CBER, contact the Office of Communication,  
25 Outreach and Development in CBER at 1-800-835-4709 or 240-402-8010 or [ocod@fda.hhs.gov](mailto:ocod@fda.hhs.gov).

26 When final, this guidance will supersede IDE Guidance Memorandum #95-2 “Implementation of  
27  
28  
29  
30  
31  
32

*Contains Nonbinding Recommendations*

*Draft – Not for Implementation*

33 the FDA/HCFA Interagency Agreement Regarding Reimbursement Categorization of  
34 Investigational Devices” issued on September 15, 1995.

35  
36  
37  
38  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60



**U.S. Department of Health and Human Services  
Food and Drug Administration  
Center for Devices and Radiological Health**

**Center for Biologics Evaluation and Research**



## **Preface**

61  
62  
63  
64  
65  
66  
67  
68  
69  
70  
71  
72  
73  
74  
75  
76  
77  
78  
79  
80  
81  
82

### **Additional Copies**

#### **CDRH**

Additional copies are available from the Internet. You may also send an e-mail request to [CDRH-Guidance@fda.hhs.gov](mailto:CDRH-Guidance@fda.hhs.gov) to receive a copy of the guidance. Please use the document number 1500074 to identify the guidance you are requesting.

#### **CBER**

Additional copies are available from the Center for Biologics Evaluation and Research (CBER), by written request, Office of Communication, Outreach, and Development (OCOD), 10903 New Hampshire Ave., Bldg. 71, Room 3128, Silver Spring, MD 20993-0002, or by calling 1-800-835-4709 or 240-402-8010, by email, [ocod@fda.hhs.gov](mailto:ocod@fda.hhs.gov) or from the Internet at <http://www.fda.gov/BiologicsBloodVaccines/GuidanceComplianceRegulatoryInformation/Guidances/default.htm>.

**Table of Contents**

83

84

85

86 I. Introduction .....5

87 II. Background.....6

88 III. FDA Interpretation of Medicare Coverage Categories A and B.....9

89 IV. Considerations When Changing from Category A to B .....10


90 V. Examples .....11

91 VI. Conclusions .....14

92 VII. Appendix A: Category Decision Flowchart.....15

93

94



95 **FDA Categorization of Investigational**  
96 **Device Exemption (IDE) Devices to**  
97 **Assist the Centers for Medicare and**  
98 **Medicaid Services (CMS) with**  
99 **Coverage Decisions**

---

100  
101  
102  
103 **Draft Guidance for Sponsors, Clinical**  
104 **Investigators, Industry, Institutional**  
105 **Review Boards and Food and Drug**  
106 **Administration Staff**

107  
108 *This draft guidance, when finalized, will represent the current thinking of the Food and Drug*  
109 *Administration (FDA or Agency) on this topic. It does not establish any rights for any person*  
110 *and is not binding on FDA or the public. You can use an alternative approach if it satisfies*  
111 *the requirements of the applicable statutes and regulations. To discuss an alternative*  
112 *approach, contact the FDA staff or Office responsible for this guidance as listed on the title*  
113 *page.*

114  
115 **I. Introduction**

116  
117 This guidance modifies the Food and Drug Administration's (FDA's or the Agency's) current  
118 policy on categorizing investigational device exemption (IDE) devices which assists the Centers  
119 for Medicare & Medicaid Services (CMS) in determining whether or not an IDE device should  
120 be covered (reimbursed) by CMS.

121  
122 On December 2, 2015, FDA's Center for Devices and Radiological Health (CDRH) and CMS's  
123 Coverage and Analysis Group (CAG) executed a Memorandum of Understanding (MOU) to

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

124 streamline and facilitate the efficient categorization of investigational medical devices in order to  
125 support CMS’s ability to make Medicare coverage (reimbursement) determinations for those  
126 investigational devices under 42 C.F.R. 405 Subpart B. The MOU noted the need for FDA and  
127 CMS to revise their shared understanding regarding categorization. This guidance document is  
128 intended to implement the MOU by further explaining the framework that FDA (both CDRH and  
129 the Center for Biologics Evaluation and Research [CBER]) intends to follow for such decisions.  
130 The MOU will take effect June 2, 2016 (6 months following signature from both FDA and CMS,  
131 as stated in the MOU). The framework in this guidance will represent the Agency’s current  
132 thinking on categorization upon publication of an FDA final guidance.

## **II. Background**

### 1995 Final Rule and FDA-HCFA Interagency Agreement

133  
134  
135  
136  
137  
138 In September 1995, the Health Care Financing Administration (now known as CMS) published a  
139 final rule and entered into an Interagency Agreement (IA) with FDA regarding reimbursement  
140 categorization of investigational devices. 60 Federal Register (FR) 48417 (September 19, 1995).  
141 The rule established that certain devices with an IDE approved by FDA (and certain services  
142 related to those devices) may be covered under Medicare, and set forth the process by which  
143 FDA would assist CMS in identifying such devices. FDA would assign a device with an FDA  
144 approved IDE to one of two categories: Experimental/Investigational (Category A) devices or  
145 Non-experimental/Investigational (Category B) devices based on the level of risk the device  
146 presented to patients. The IA set forth criteria, agreed upon by CMS and FDA, that FDA would  
147 use to categorize devices. The categorization would then be used by CMS as part of its  
148 determination of whether or not items and services met the requirements for Medicare coverage  
149 under Section 1862(a)(1)(A) of the Social Security Act (the “reasonable and necessary” clause).  
150 That is, to be eligible to be covered (e.g., to have a benefit category determination) under  
151 Medicare, the device must be reasonable and necessary for the diagnosis or treatment of an  
152 illness or injury, or to improve the functioning of a malformed body member.<sup>1</sup>

153  
154  
155  
156 Under the 1995 CMS final rule, Category A devices were devices believed to be in class III for  
157 which “absolute risk” of the device type had not yet been established. That is, initial questions  
158 of safety and effectiveness had not been resolved and FDA was unsure whether the device type  
159 could be safe and effective. The IA contained two sub-categories which provided criteria  
160 indicating that a given device met this standard and should be placed into Category A: those  
161 devices for which no marketing application had been approved through the premarket approval  
162 (PMA) process for any indication for use and devices that would otherwise be a Category B, but

---

<sup>1</sup> Implementation of the FDA/HCFA Interagency Agreement Regarding Reimbursement Categorization of Investigational Devices, Att. A Interagency Agreement, Att. B Criteria for Categorization of Investigational Devices, & Att. C List. #D95-2 (IDE Guidance Memorandum #95-2, Sept. 15, 1995).

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

163 had undergone significant modification for a new indication or use. An example of a significant  
164 modification may be the addition of a drug onto a legally marketed device.

165  
166 Under the 1995 CMS final rule, Category B devices were those devices believed to be in Class I  
167 or II, or devices believed to be in Class III for which the incremental risk was the primary risk in  
168 question. That is, underlying questions of safety and effectiveness of that device type had been  
169 resolved or it was known that the device type could be safe and effective because, for example,  
170 other manufacturers had obtained FDA approval for that device type. The IA identified six sub-  
171 categories of investigational devices that were of a device type for which the underlying  
172 questions of safety and effectiveness had been resolved and thus should be placed in Category B.  
173 Under the IA, Category B devices included those that were under investigation to demonstrate  
174 substantial equivalence to a predicate device (legally marketed device) through the 510(k)  
175 process or devices comparable to a PMA-approved device. Category B also included situations  
176 in which it was known that the device type could be safe and effective because, for example,  
177 other manufacturers had obtained FDA approval for that device type. Several examples of  
178 Category A and B devices can be found later in this document.

179  
180 Importantly, CMS and FDA both recognized that experience in categorizing devices might  
181 require changes to the Interagency Agreement.<sup>2</sup>

#### 2013 Amendment to 42 CFR 405 Subpart B

182  
183  
184  
185 In 2013, CMS published a final rule in the Federal Register (FR), 78 FR 74230, 74809 (Dec. 10,  
186 2013), that, among other things, modified the definitions for Category A and Category B. These  
187 definitions can be found in the Code of Federal Regulations (CFR) at 42 CFR 405.201:

188  
189 *Category A (Experimental)*  
190 42 CFR 405.201(b): "...a device for which 'absolute risk' of the device type has not been  
191 established (that is, initial questions of safety and effectiveness have not been resolved) and the  
192 FDA is unsure whether the device type can be safe and effective."

193  
194 *Category B (Nonexperimental/investigational)*  
195 42 CFR 405.201(b): "...a device for which the incremental risk is the primary risk in question  
196 (that is, initial questions of safety and effectiveness of that device type have been resolved), or it  
197 is known that the device type can be safe and effective because, for example, other  
198 manufacturers have obtained FDA premarket approval or clearance for that device type."

199  
200 CMS uses FDA's categorization determination in evaluating whether or not an IDE device  
201 receives Medicare coverage. Medicare may make payment for an investigational device and  
202 routine care items and services furnished in an FDA-approved Category B  
203 (Nonexperimental/Investigational) IDE study if CMS (or its designated entity) determines prior  
204 to the submission of the first related claim that the Medicare coverage IDE study criteria in 42

---

<sup>2</sup> The Interagency Agreement was published as an addendum to the final rule in 1995. The FR noted that: "As experience is gained in the categorization process, this addendum may be modified." 60 FR at 48419.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

205 CFR 405.212 are met.<sup>3</sup> Medicare may cover only routine care items and services furnished in an  
206 FDA-approved Category A (Experimental) IDE study, but not the device itself if CMS (or its  
207 designated entity) determines that Medicare coverage IDE study criteria in 42 CFR 405.212 are  
208 met.<sup>4</sup> In other words, Medicare cannot cover device expenses for studies that FDA has  
209 categorized as Category A (Experimental).

210

211

#### Reasons for Modification of the Previous FDA Policy

212

213 In the more than twenty years since the IA was signed, FDA has received a number of IDEs  
214 which do not easily fit into any of the eight sub-categories identified in the IA.

215

216 In 2013, FDA published a final guidance document entitled “Investigational Device Exemptions  
217 (IDEs) for Early Feasibility Medical Device Clinical Studies, Including Certain First in Human  
218 (FIH) Studies.” This document provides guidance on the development and review of IDE  
219 applications for early feasibility studies (EFS) of significant risk devices. EFS are feasibility  
220 studies that are very small in size and allow for early clinical evaluation of devices that may not  
221 be a final design. They are intended to provide proof of principle and initial clinical study data.  
222 Traditional feasibility studies, on the other hand, are completed with a device design that is near-  
223 final or final and are commonly used to capture preliminary safety and effectiveness information  
224 which may be used to inform a pivotal study design. They are typically larger than EFS. The  
225 general term “feasibility studies” may refer to EFS or traditional feasibility studies. Pivotal  
226 studies are clinical investigations designed to collect definitive evidence of the safety and  
227 effectiveness of a device for a specified intended use, typically in a statistically justified number  
228 of subjects. The previous FDA policy regarding reimbursement categorization did not adequately  
229 articulate categorization criteria that are relevant to certain feasibility studies, particularly those  
230 for devices similar to approved devices but with modifications which raise significant new safety  
231 questions. As a result of this and the recent increase in EFS submissions subsequent to the  
232 publication of the guidance document referenced above, FDA has determined that additional  
233 clarification of these categorization criteria is warranted. It is important to note that the CMS  
234 category designation is made independent of study type and instead is based on the criteria  
235 described in this document.

236

237 In addition to the above consideration, there are situations when adequate data are provided to  
238 resolve initial questions of safety and effectiveness (e.g., data from a feasibility study becomes  
239 available) and, therefore, it is appropriate to change the device category for subsequent studies of  
240 the same device from Category A to Category B. In these circumstances, a device that had  
241 previously been categorized as experimental could now be considered  
242 nonexperimental/investigational. However, the IA did not describe a pathway for changing  
243 categorization from Category A to Category B when approving subsequent studies for the same  
244 device. In order to outline a mechanism to revisit the categorization of IDE devices when new  
245 information is gathered, the previous FDA policy for CMS categorization of IDE devices is  
246 being modified.

---

<sup>3</sup> 42 CFR 405.211(b)

<sup>4</sup> 42 CFR 405.211(a)

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

247  
248 Lastly, in its changes to the regulations (42 CFR 405 Subpart B), effective January 1, 2015, CMS  
249 added criteria for coverage of IDE studies and changed from local Medicare Administrative  
250 Contractor (MAC) review and approval of IDE studies to a centralized review and approval of  
251 IDE studies. The change to a centralized IDE review further reinforced the need for CMS and  
252 FDA to revisit the policy that FDA used to categorize IDE devices. CMS and FDA recognized  
253 the necessity to revise their shared understanding regarding the categorization of IDE devices to  
254 help ensure that devices will not be precluded from reimbursement due to an inappropriate  
255 reimbursement categorization determination. Rather than amending their 1995 IA, FDA and  
256 CMS entered into an MOU on December 2, 2015. It becomes effective on June 2, 2016. The  
257 policies and framework in this guidance will represent the Agency’s current thinking on  
258 categorization upon publication of a final guidance document.  
259  
260

### 261 **III. FDA Interpretation of Medicare Coverage Categories A** 262 **and B**

263  
264 After receipt of an IDE application, FDA will determine whether the sponsor has provided  
265 enough information to support initiation of the clinical study. An IDE application is “approved”  
266 or “approved with conditions” if FDA has determined that the sponsor has provided adequate  
267 data to support initiation of a human clinical study, no subject protection concerns preclude  
268 initiation of the investigation, and the benefit-risk profile is sufficiently favorable to justify  
269 enrollment.<sup>5</sup> FDA intends to use the criteria described below to assign a device to a CMS  
270 Category A or B when the IDE is approved or approved with conditions. Please refer to  
271 Appendix A for a flowchart depicting the decision making process.  
272

#### 273 **Category A: Experimental**

274  
275 42 CFR 405.201(b): “...a device for which ‘absolute risk’ of the device types has not been  
276 established (that is, initial questions of safety and effectiveness have not been resolved) and the  
277 FDA is unsure whether the device type can be safe and effective.” FDA intends to consider a  
278 device to be in Category A if one or more of the following criteria are met:  
279

- 280 • No PMA approval, 510(k) clearance or *de novo* request has been granted for the proposed  
281 device or similar devices, and non-clinical and/or clinical data on the proposed device do  
282 not resolve initial questions of safety and effectiveness.  
283
- 284 • The proposed device has different characteristics compared to a legally marketed device;  
285 and information related to the marketed device does not resolve initial questions of safety

---

<sup>5</sup> For more information on how IDE Decisions are made please refer to the FDA Guidance document “[FDA Decisions for Investigational Device Exemption Clinical Investigations](#).”

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

286 and effectiveness for the proposed device. Available non-clinical and/or clinical data on  
287 the proposed device also do not resolve these questions.  
288

- 289 • The proposed device is being studied for a new indication or new intended use for which  
290 information from the proposed or similar device related to the previous indication does  
291 not resolve initial questions of safety and effectiveness. Available non-clinical and/or  
292 clinical data on the proposed device relative to the new indication or intended use also do  
293 not resolve these questions.

294

#### **Category B: Nonexperimental/Investigational**

295

296  
297 42 CFR 405.201(b): "...a device for which the incremental risk is the primary risk in question  
298 (that is, initial questions of safety and effectiveness of that device type have been resolved), or it  
299 is known that the device type can be safe and effective because, for example, other  
300 manufacturers have obtained FDA premarket approval or clearance for that device type.”

301

302 FDA intends to consider a device to be in Category B if one or more of the following criteria are  
303 met:  
304

305

- 306 • No PMA approval, 510(k) clearance or *de novo* request has been granted for the proposed  
307 device or similar devices; however, available clinical data (e.g., feasibility study data)  
308 and/or non-clinical data for the proposed device or a similar device resolve the initial  
309 questions of safety and effectiveness.

310

- 311 • The proposed device has similar characteristics compared to a legally marketed device,  
312 and information related to the marketed device resolves the initial questions of safety and  
313 effectiveness for the proposed device. Additional non-clinical and/or clinical data on the  
314 proposed device may have been used in conjunction with the leveraged<sup>6</sup> information to  
315 resolve these questions.

316

- 317 • The proposed device is being studied for a new indication or new intended use; however,  
318 information from the proposed or similar device related to the previous indication  
319 resolves the initial questions of safety and effectiveness. Additional non-clinical and/or  
320 clinical data on the proposed device may have been used in conjunction with the  
321 leveraged information to resolve these questions.

322

## **IV. Considerations When Changing from Category A to B**

323

---

<sup>6</sup> For purposes of this draft guidance, the term “leveraged” means that data from the legally marketed device are relevant to the proposed device, were determined to be valid scientific evidence, and may be used to help resolve initial questions of safety and effectiveness.

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

324 As mentioned previously in this document, there are situations in which non-clinical and/or  
325 clinical evaluations provide adequate data to resolve initial questions of safety and effectiveness  
326 and, therefore, it is appropriate to change the device category for subsequent studies of the same  
327 device from Category A to Category B. For example, a categorization change may be justified  
328 when a completed study, in which the device was designated as Category A, has resulted in  
329 clinical data that resolve the initial questions of safety and effectiveness. In this case, the device  
330 may then be designated as Category B in the subsequent study.

331  
332 Another situation where a category change may be warranted is when an IDE study receives a  
333 staged approval or staged approval with conditions.<sup>7</sup> In a staged approval, FDA may grant IDE  
334 approval or approval with conditions for a portion of the intended study cohort, enabling certain  
335 outstanding questions to be answered concurrently with enrollment in this cohort. The sponsor  
336 will be permitted to expand enrollment once an IDE supplement containing the necessary  
337 additional information is submitted to FDA and found to be acceptable. In some cases, the  
338 purpose of the initial stage of the clinical study is to resolve initial questions of safety and  
339 effectiveness. In this situation the device will be designated as Category A for the initial stage. If  
340 adequate data are gathered from the initial stage of the study such that the initial questions of  
341 safety and effectiveness have been resolved and the sponsor has been granted expanded  
342 enrollment, the category may be changed from Category A to Category B for the device in the  
343 expanded study.

344  
345 FDA will evaluate whether adequate data are present to resolve the initial questions of safety and  
346 effectiveness and a categorization decision will be made upon study approval (for a new study),  
347 study expansion (for a staged study), or submission of a request to change the category. A  
348 request to change the category should be submitted as an IDE supplement. The categorization  
349 decision will be included in either the IDE approval letter to the sponsor or a letter to the sponsor  
350 in response to a request for category change.

351

## 352 **V. Examples**

353

### 354 **Category A: Experimental**

355 The list below provides examples of when a Category A determination may be appropriate, but it  
356 does not represent an exhaustive list of when a device should be classified as Category A.

357

- 358 • A device is completely novel and has no, or limited, previous human use and there are  
359 initial questions of safety and effectiveness. There is adequate non-clinical information to  
360 support initiation of an early feasibility study that will provide data to inform potential  
361 device design or procedural improvements.
- 362  
363 • A drug is added to a previously approved or cleared device. While substantial  
364 information is known about the previously approved or cleared device, the addition of a

---

<sup>7</sup> See "[FDA Decisions for Investigational Device Exemption Clinical Investigations.](#)"

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

365 drug has resulted in initial questions of safety and effectiveness that have not yet been  
366 resolved.

367

- 368 • An already approved or cleared device is being evaluated for a new intended use or  
369 indication wherein the device will be placed in a different anatomical location. The  
370 device's technology is unchanged from what was initially approved; however, it is  
371 uncertain as to whether the device can be safely placed in the new anatomical location  
372 and whether the device can also be effective in the new anatomical location. Therefore,  
373 there are inadequate data to resolve the initial questions of safety and effectiveness  
374 relative to the new intended use or indication.

375

- 376 • The initial question of safety has been answered with the submission of non-clinical  
377 and/or clinical data. There is inadequate evidence to resolve initial questions related to  
378 effectiveness; however, the benefit-risk profile supports initiation of a pivotal study.

379

#### **Category B: Nonexperimental/Investigational**

381 The list below provides examples of when a Category B determination may be appropriate, but it  
382 does not represent an exhaustive list of when a device should be classified as Category B.

383

- 384 • The insertion system of an approved device has been modified to improve ease of use for  
385 the clinician. Non-clinical test data resolved initial questions of safety and effectiveness  
386 related to this change; however, confirmatory clinical information about the device  
387 performance is required due to the inherent differences between the non-clinical test  
388 environment and the clinical setting. (The non-clinical data and a benefit-risk assessment  
389 support initiation of a small feasibility study to resolve this incremental risk and inform  
390 the final device design.)

391

- 392 • Adequate data have been gathered from non-clinical testing and the clinical results of a  
393 feasibility study such that initial questions of safety and effectiveness have been resolved.  
394 A pivotal study will be initiated to provide the primary clinical evidence for the safety  
395 and effectiveness of the device in support of a future marketing application.

396

- 397 • A range of device sizes will be included in a clinical study, but data that resolve initial  
398 questions of safety and effectiveness have been received on only a subset of the sizes. It  
399 is anticipated that the data for the other sizes will also resolve initial questions; therefore,  
400 the study will be staged. In this case, the study will start with the initially approved  
401 device sizes while additional supportive information is collected on the remaining device  
402 sizes. Because the initial questions of safety and effectiveness have been resolved for the  
403 initial stage and will be resolved for the additional device sizes prior to expansion of the  
404 study, the devices in both the initial stage and expanded study will be designated  
405 Category B.

406

- 407 • A new device will be studied for an indication for which substantial safety and  
408 effectiveness information exists from other similar device(s) of the same type that are  
409 used for the same or similar indication. Non-clinical test data that have been provided

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

410 can answer initial questions regarding the anticipated device performance relative to this  
411 indication. Because the initial questions of safety and effectiveness have been resolved, a  
412 pivotal study to evaluate this new device will be designated Category B.

- 413
- 414 • A modification has been made to an approved device in order to improve its  
415 performance. Non-clinical and clinical data available from the previous version of the  
416 device along with additional testing on the modified device resolved initial questions of  
417 safety and effectiveness. The purpose of the study will be to gather further data regarding  
418 device performance for this modified version of the device.
- 419
- 420 • New device sizes will be added to a product matrix for an approved device. Initial  
421 questions of safety and effectiveness have been resolved based on experience with the  
422 approved device, and it is generally understood how the new device sizes will perform.  
423 The new device sizes will be studied such that statistical information on safety and  
424 effectiveness relevant to these sizes can be gathered.
- 425
- 426 • An approved device will be evaluated in a new patient population. Non-clinical and  
427 clinical data from use in the previous patient population resolved initial questions of  
428 safety and effectiveness for the new patient population. The new study to be conducted  
429 will provide further data regarding device performance for this new patient population.
- 430
- 431 • An approved device will be evaluated for a new indication. Data exist on the approved  
432 device for another similar indication, and non-clinical data have also been supplied such  
433 that the initial questions of safety and effectiveness related to the new indication have  
434 been resolved. The new study to be conducted will provide further data regarding device  
435 performance for this new indication.
- 436
- 437 • A new device will be studied for an indication in which there are no other devices of a  
438 similar type. However, the non-clinical test data supplied are robust and resolve the initial  
439 questions of safety and effectiveness. The study to be conducted will provide further data  
440 regarding device performance for this indication.

#### **Change from Category A to Category B**

443 If the device was previously designated as Category A, but the initial questions of safety and  
444 effectiveness of the device have since been resolved, it may be appropriate to change the  
445 Category from A to B. The list below provides examples of when a change from Category A to  
446 Category B may be appropriate, but it does not represent an exhaustive list of when a device may  
447 change from Category A to Category B.

- 448
- 449 • A novel insertion procedure will be used to place an already approved or cleared device  
450 and there are initial questions of safety and effectiveness regarding the novel insertion  
451 procedure that have not been resolved. In this case, these questions of safety and  
452 effectiveness may be answered in a short time frame with a limited number of subjects in  
453 the context of a larger clinical study. Therefore, the device will be evaluated in a staged  
454 clinical study where the first stage falls under Category A. If the initial questions of

## *Contains Nonbinding Recommendations*

### *Draft – Not for Implementation*

455 safety and effectiveness are resolved and the study continues, the device may be re-  
456 categorized to Category B.

457

458 • Adequate data have been gathered on a device from non-clinical testing, the completion  
459 of an early feasibility study within the United States (US), as well as a small non-US  
460 clinical study such that initial questions of safety and effectiveness have been resolved.  
461 Additional data are needed to help inform a pivotal study design; therefore, a traditional  
462 feasibility study will be initiated. Although the EFS was originally designated as  
463 Category A, adequate data as described above have since been gathered to support a  
464 change to Category B for the traditional feasibility study.

465

466 • A device is currently being evaluated in a clinical study and has been designated  
467 Category A. While the study is being conducted, clinical study results for comparable  
468 products became available which resolve initial questions of safety and effectiveness for  
469 the device. This information will be used to support a categorization change from  
470 Category A to Category B for the device evaluated in the ongoing clinical study.

471

## **VI. Conclusions**

472

473


474 FDA categorizes IDE devices based on whether available data demonstrate that initial questions  
475 of safety and effectiveness have been resolved. This guidance document describes the criteria  
476 that will be used to help determine the appropriate category for a device to be studied. This  
477 guidance document also describes when it is appropriate to change the device category from  
478 Category A to Category B. The categorization of IDE devices is used by CMS as part of its  
479 determination of which devices meet the requirements for Medicare coverage under Section  
480 1862 (a)(1)(A) of the Social Security Act (the “reasonable and necessary” clause). IDE device  
481 categorization is only part of the information used to determine coverage by CMS. Please refer to  
482 the website “[Medicare Coverage Related to Investigational Device Exemption \(IDE\) Studies](#)” for  
483 guidance on requesting coverage and for contact information.

484

485

486

## VII. Appendix A: Category Decision Flowchart



\* S/E = Safety and Effectiveness

487