		TH AND HUMAN SERVICES G ADMINISTRATION			
DISTRICT ADDRESS AND PHON		DATE(S) OF INSPECTION			
158-15 Libert	-	10/29/2013 - 11/06/ FEINUMBER	2013*		
Jamaica, NY (718) 340-700		1316297			
,	ormation: www.fda.gov/oc/indu				
TO: Bruce B	. Whitcavitch, Jr., Vice-Pres	ident of Manufacturing/Operation STREET ADDRESS	ns		
	national, Inc.	141 Sal Landrio Dr			
I I I I I I I I I I I I I I I I I I I	idefendi, inc.	Crossroads Business Park			
CITY, STATE, ZIP CODE, COUNT		TYPE ESTABLISHMENT INSPECTED			
Johnstown, NY	12095-3835	Medical Device Manufacturer			
observations, and do observation, or have action with the FDA questions, please con	not represent a final Agency determination regatimplemented, or plan to implement, corrective a representative(s) during the inspection or submitact FDA at the phone number and address about the phone for this Form FDA-483 are not an exhibit of the phone for the phone	austive listing of objectionable conditions. Unde	arding an sthe objection or u have any		
firm is responsible requirements.	for conducting internal self-audits to ident	ify and correct any and all violations of the qua	lity system		
DURING AN INSPEC	CTION OF YOUR FIRM I OBSERVED:				
OBSERVATION	1				
OBOLKVATION	•				
Process control probeen established.	ocedures that describe any process controls	necessary to ensure conformance to specification	ons have not		
Specifically, you as (b) (4)	re fielding complaints and remanufacturing , without establishe	the (b) (4) product, a product, a production & process control procedures for t			
#(b) (4) dated 1 Your firm disassen	For example, your firm has received complaint #(b) (4) dated 8/6/13, complaint #(b) (4) dated 9/16/13 and complaint #(b) (4) serial numbers (b) (4) Your firm disassembled and/or repaired all three probes and sent them back to the customers for use without the use of established production & process control procedures.				
_					
OBSERVATION	2				
Procedures for desi	ign change have not been established.				
mechanical require Control inspection from having an act specifications in th implemented via E	ment of bond strength on the Extension Se manufacturing procedure, SOP-457 'Exten ion limit of (b) (4) lbs. to and action limit of (b) e Extension Sets design file, Project (b) (4) CO (b) (4), through the entire design review	o/15/2006 changed the production specification ts, Project (b) (4) , from (b) (4) 1 sion Set Bond Strength' was updated at Rev.6 o (4) lbs. However, your firm did not update the for this attribute or carry this specification process to ensure the specification change does nge in specification was not adequately docume	bs. The Quality n 2/3/2011 design ion change, not adversely		
	EMPLOYEE(S) SIGNATURE		DATE ISSUED		
SEE REVERSE OF THIS PAGE	Catherine M. Beer, Investig	ator	11/06/2013		
FORM FDA 483 (09/08)	PREVIOUS EDITION OBSOLETE INSPH	CCTIONAL OBSERVATIONS	PAGE 1 OF 3 PAGES		

DEPARTMENT OF HEALTH AND HUMAN SERVICES FOOD AND DRUG ADMINISTRATION					
DISTRICT ADDRESS AND PHONE NUMBER		DATE(S) OF INSPECTION			
158-15 Liberty Ave.		10/29/2013 - 11/06/2013*			
Jamaica, NY 11433		FEI NUMBER			
(718) 340-7000 Fax:(718) 662-5661		1316297			
Industry Information: www.fda.gov/oc/industry					
NAME AND TITLE OF INDIVIDUAL TO WHOM REPORT ISSUED					
TO: Bruce B. Whitcavitch, Jr., Vice-President of Manufacturing/Operations					
FIRM NAME	STREET ADDRESS				
Epimed International, Inc.	141 Sal Landrio Dr				
	Crossroads Business Park				
CITY, STATE, ZIP CODE, COUNTRY	TYPE ESTABLISHMENT INSPECTED				
Johnstown, NY 12095-3835	Medical Device Manufacturer				

## **OBSERVATION 3**

Procedures for design validation have not been established.

Specifically, design validation procedure SOP-705 Rev. 5 'Design Validation' does not provide adequate instruction to ensure all deviations from acceptance criteria are identified and corrected prior to the approval of the validation data.

For Example, 'Design Validation (b) (4) Catheter' for Project No.(b) (4) dated 8/16/2013 was completed with a missing datapoint for 'Shear Resistance' of the catheter. The acceptance criteria of the validation required an average score of (b) (d) must be reviewed to determine if the failure affects the safety or effectiveness of the device. The missing datapoint for 'Shear Resistance' of the catheter was not identified during validation data review, complilation of the summary report or conclusion. The validation was approved without the deviation identified or corrected and the design validation did not meet acceptance criteria.

Additionally, the design review conducted 10/1/2013 for 'Design Validation Close Out', after the validation was completed did not identify the lacking data/deviation in the design validation.

SEE REVERSE OF THIS PAGE	Catherine M. Beer,	Investigator	11/06/2013
EI	EMPLOYEE(S) SIGNATURE		DATE ISSUED

## DEPARTMENT OF HEALTH AND HUMAN SERVICES FOOD AND DRUG ADMINISTRATION DISTRICT ADDRESS AND PHONE NUMBER DATE(S) OF INSPECTION 158-15 Liberty Ave. 10/29/2013 - 11/06/2013\* FEI NUMBER Jamaica, NY 11433 1316297 (718) 340-7000 Fax: (718) 662-5661 Industry Information: www.fda.gov/oc/industry NAME AND TITLE OF INDIVIDUAL TO WHOM REPORT ISSUED Bruce B. Whitcavitch, Jr., Vice-President of Manufacturing/Operations Epimed International, Inc. 141 Sal Landrio Dr Crossroads Business Park CITY, STATE, ZIP CODE, COUNTRY TYPE ESTABLISHMENT INSPECTED Johnstown, NY 12095-3835 Medical Device Manufacturer **Observation Annotations** Observation 1: Promised to correct. Observation 2: Promised to correct. Observation 3: Promised to correct.

## \* DATES OF INSPECTION:

10/29/2013(Tue), 10/30/2013(Wed), 10/31/2013(Thu), 11/01/2013(Fri), 11/04/2013(Mon), 11/06/2013(Wed)

	TANGER CONTACT OR CERTIFICATION OF	
SEE REVERSE OF THIS PAGE	Catherine M. Beer, Investigator	11/06/2013
	EMPLOYEE(S) SIGNATURE	DATE ISSUED