


Medical Device Recall 2013 Update

Ron Brown
Chief for the Recall Branch
Division of Risk Management Operations
Office of Compliance
Center for Devices and Radiological Health
Food and Drug Administration








Probable Causes of Recent Increases

- Elimination of backlog within FDA
- Increased vigilance of device manufacturers
- Increased reporting by Industry
- Increased number of devices and manufacturers



Recall Causes in FY 2012





Recall Challenges

- Understanding the Risk
- Adequate Notification
- Recall Cause Analysis
- Corrective and Preventive Actions
- Closing a Recall
- Terminating a Recall



What is a recall?

- Firm's removal or correction of a marketed product in violation of the F, D, & C Act and against which FDA would initiate legal action
 - Voluntary action taken by a firm when they determine a device is misbranded § 502 or adulterated § 501 of the Act
 - Effective method to remove or correct consumer products from the market place
 - Alternative to FDA initiated court action for removing violative products from the market (seizure) or import detention



Classes of Recalls

- O Class I- Situation in which there is a reasonable probability that the use of, or exposure to, a violative product will cause serious adverse health consequences or death
- o Class II- Situation in which use of, or exposure to, a violative product may cause temporary or medically reversible adverse health consequences or where the probability of serious adverse health consequences is remote
- O Class III Situation in which use of, or exposure to, a violative product is not likely to cause adverse health consequences.



Recall Regulations

- 21 CFR 7 (7.40-7.59 Recalls)
 - How to conduct a recall
 - Guidance for firms (voluntary)
 - Requirements for FDA
- 21 CFR 806 (Medical Device Reports of Removals and Corrections)
 - Reporting requirement for medical device manufacturers & importers
 - for situations with a risk to health (class 1 & 2 recalls)
- 21 CFR 810 (Mandatory Medical Device Recalls)
 - Rarely used



Related Regulations

- o 21 CFR 820 Quality System Regulation
 - GMPs
 - Design control
 - Product & Process changes
 - Validation of device and process changes
 - Acceptance activities
 - CAPA
 - Complaint files
 - * etc
- 21 CFR 803 Medical Device Reporting
- 21 CFR 807.81(a)(3) Pre-Market Notification when changes are introduced



Ways CDRH May Become Aware of a Recall Situation

- 806 Correction and Removal Reports
- FDA Inspections
 - Directed
 - Routine
- Consumer Complaints
- Adverse Event Reports
 - Medwatch Report
 - MDRs (Medical Device Report), OSB

- Information received by a firm, repackager or distributor
- Competitors
- Internal Device Reviews
- MedSun Reports
- MRA and other Country Agreements
- Radiation Product Reports

List is not all inclusive




Triggering Events

- o 21 CFR 806.10(a)
 - Each device manufacturer or importer shall submit a written report to your FDA district office of any correction or removal of a device initiated by such manufacturer or importer if the correction or removal was initiated:
 - 1. To reduce a risk to health posed by the device; or
 - To remedy a violation of the act caused by the device which may present a risk to health
- o 21 CFR 806.10(b)
 - The manufacturer or importer shall submit any required report within 10-working (business) days of initiating such correction or removal.




21 CFR Part 7.42: A Firm's Recall Strategy

- The recall strategy will include the following elements:
 - Depth level in the distribution chain
 - Public Warning purpose is to alert the public that the product being recalled presents a serious hazard to health
 - Effectiveness Checks verifies that all consignees at the recall depth specified have received notification and have taken appropriate action
- The recall strategy will specify the method(s) to be used for and the level of effectiveness checks that will be conducted



A Risk Assessment Should Be Conducted:

- O Prior to
 - Stock Recoveries
 - Safety Alerts
 - Market Withdrawals
 - Recalls




Other FDA/CDRH information

- Part 806 Reports of Corrections and Removals
- http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cf cfr/CFRSearch.cfm?CFRPart=806
- Customer Notification Letter & Press Release template and other important links
- http://www.fda.gov/Safety/Recalls/IndustryGuidance/ default.htm
- CDRHLearn modules
- http://www.fda.gov/Training/CDRHLearn/ucm162015.
 htm



Other FDA/CDRH information

- Distinguishing Medical Device Recalls from Product Enhancements – Draft Guidance
- http://www.fda.gov/MedicalDevices/DeviceRegulation andGuidance/GuidanceDocuments/ucm340518.htm
- Guidance for Industry: Product Recalls, Including Removals and Corrections
- o http://www.fda.gov/Safety/Recalls/IndustryGuidance/ucm129259.htm



Other FDA/CDRH information

- News <u>http://www.fda.gov/MedicalDevices/NewsEvents/default.htm</u>
- Device approvals -<u>http://www.fda.gov/MedicalDevices/ProductsandMedicalProcedures/DeviceApprovalsandClearances/Recently-ApprovedDevices/default.htm</u>
- Compliance actions
 http://www.fda.gov/MedicalDevices/DeviceRegulation
 andGuidance/ComplianceActivities/default.htm

A to Z Index | Follow FDA | FDA Voice Blog

SEARCH

Most Popular Searches

Home

Food

Drugs

Medical Devices

Radiation-Emitting Products

Vaccines, Blood & Biologics

Animal & Veterinary

Cosmetics

Tobacco Products

Medical Devices

Home Medical Devices



Medical Devices Topics

Products and Medical Procedures

Approvals & Clearances, Home Use, Surgical, Implants & Prosthetics. In Vitro Diagnostics, more...

Medical Device Safety

Alerts & Notices, Recalls, Report a Problem, MedSun, Emergency Situations

Device Advice: Comprehensive Regulatory Assistance How to Market a Device, Postmarket Requirements, Science and Research (Medical Devices)

Chemistry & Materials Science, Solid & Fluid Mechanics, Imaging & Applied Mathematics, Electrical & Software Engineering, more...

News & Events (Medical Devices)

Medical Device News, Videos, Workshops & Meetings

Resources for You (Medical Devices)

Consumers, Health Care Providers, Regulated Industry

Spotlight

- FDA announces public-private partnership to develop regulatory science that will speed patient access to new medical device technologies
- Improvements in Device Review: Results of CDRH's Plan of Action for Premarket Review of Devices
- National Medical Device Postmarket Surveillance Plan
- CDRH Mission, Vision and Shared Values
- CDRH 2013 Strategic Priorities

Recalls & Alerts

- FDA warns against improper advertising, promotion of lasers intended for LASIK corrective eye surgery
- Narrowed Indications for Use for











Home

Food

Drugs

Medical Devices

Vaccines, Blood & Biologics

Animal & Veterinary

Cosmetics

Radiation-Emitting Products

Tobacco Products

Medical & Radiation Emitting Device Recalls

FDA Home Medical Devices Databases



Page Last Updated: 02/12/2013

Note: If you need help accessing information in different file formats, see Instructions for Downloading Viewers and Players.

Go to Simple Search



FDA Basics

FOIA

Records per Report Page

No Fear Act

Site Map

Clear

Transparency

Website Policies

10



A to Z Index | Follow FDA | FDA Voice Blog SEARCH

Home

Food

Drugs

Medical Devices

Vaccines, Blood & Biologics

Animal & Veterinary

Cosmetics

Radiation-Emitting Products

Most Popular Searches

Tobacco Products

MAUDE - Manufacturer and User Facility Device Experience







- FDA Home Medical Devices Databases
 - MAUDE data represents reports of adverse events involving medical devices. The data consists of voluntary reports since June 1993, user facility reports since 1991, distributor reports since 1993, and manufacturer reports since August 1996. MAUDE may not include reports made according to exemptions, variances, or alternative reporting requirements granted under 21 CFR 803.19.
 - The on-line search allows you to search CDRH database information on medical devices which may have malfunctioned or caused a death or serious injury. MAUDE is scheduled to be updated monthly and the search page reflects the date of the most recent update. FDA seeks to include all reports received prior to the update. However, the inclusion of some reports may be delayed by technical or clerical difficulties.
 - MAUDE data is not intended to be used either to evaluate rates of adverse events or to compare adverse event occurrence rates across devices. Please be aware that reports regarding device trade names may have been submitted under different manufacturer names. Searches only retrieve records that contain the search term(s) provided by the requester.



The Freedom of Information Act (FOIA) is found in Title 5 of the United States Code, Section 552. FOIA generally provides that any person has the right to request access to federal agency records or information except to the extent the records are protected from disclosure by any of nine exemptions

contained in the law or by one of three anexial law enforcement record evaluations. FOIA also condite to the release of information in the reports found in



Thank You

Questions?